

Certificate of Compliance

Certificate: 1804189

Master Contract: 235496

Project: 1804189


Date Issued: June 21, 2006

Issued to: EUROIMMUN Medizinische Labordiagnostika AG
Seekamp 31
Lübeck, 23560
GERMANY


The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by:


Carole Lemay

Authorized by:

 for:
Helene Vaillancourt
Operations Manager

PRODUCTS

CLASS 8721 05 – LABORATORY EQUIPMENT Electrical

CLASS 8721 85 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - CERTIFIED TO U.S. STANDARDS

- Stratec Automatic analyzer for invitro diagnostic, model EUROIMMUN Analyzer I, type "6398"; Rated 115 V~, 60 Hz, 4A / 230 V~, 50 Hz, 2A

Note: The above model is Cord connected, Equipment Class 1, Pollution Degree 2, Installation Category II.

MARKINGS

Listee's name and/or CSA file number 235496, model designation, complete electrical rating in volts, hertz, hp, amps, serial number or date coding, and the CSA Mark appear in a permanent manner on each unit.

The cCSAus monogram may appear next to the CSA Mark