

EUROIMMUN Medizinische Labordiagnostika AG

EUROIMMUN
Medizinische Labordiagnostika AG
Seekamp 31
23560 Lübeck

☎ 0451-2032-0
✉ www.euroimmun.de/kontakt
🌐 www.euroimmun.de

22. Dezember 2025

Statement

EUROIMMUN Medizinische Labordiagnostika AG declares that all *IVDD-Legacy Devices* meet the demands of Directive 98/79/EC (IVDD) based on conformity assessment procedure according to annex III. This remains valid on basis of article 110 (3) of the Regulation (EC) 2017/746 (IVDR).

However, the requirements of this Regulation relating to post-market surveillance, market surveillance, vigilance, registration of economic operators and of devices shall apply and replace the corresponding requirements in Directive 98/79/EC.

IVDD-Legacy devices are exempted from the Regulation (EC) No 1272/2008 (CLP) as stated in *article 1*, section 5 (d).

Therefore, classification and labelling obligations according to CLP and subsequently the requirements of regulation (EC) No 1907/2006 (REACH) regarding safety data sheets are not applicable.

Head of Division Quality Management and Regulatory Affairs